



Certificate of Analysis

Standard Reference Material[®] 1616b

Sulfur in Kerosene (Low Level)

This Standard Reference Material (SRM) is intended for use in the evaluation of methods and the calibration of instruments used in the determination of total sulfur in fuel oils or materials of a similar matrix. SRM 1616b consists of a special low sulfur kerosene (No. 1-K) suitable for use in non flue-connected burner appliances and for use in wick-fed illuminating lamps, as described in ASTM D 3699-92 Specification for Kerosene [1]. A unit of SRM 1616b consists of 100 mL of the low sulfur kerosene in an amber glass bottle.

Certified Value: The measurand is the mass fraction for sulfur listed in Table 1. The certified value is metrologically traceable to the SI derived units for mass fraction (expressed as milligrams per kilogram). The certified sulfur content is based on analyses by isotope dilution thermal ionization mass spectrometry (ID-TIMS) [2]. A NIST certified value is a value for which NIST has the highest confidence in its accuracy in that all known or suspected sources of bias have been investigated or taken into account. Homogeneity testing was performed using X-ray fluorescence spectrometry (XRF). The expanded uncertainty for the certified value for sulfur is calculated as a 95 % confidence interval where $U = ku_c$. The quantity u_c is intended to represent, at the level of one standard deviation, the combined standard uncertainty calculated according to the ISO/JCGM Guides [4]. The coverage factor, $k = 2.23$, corresponds to a t factor obtained from the t -distribution for approximately 10.3 degrees of freedom.

Table 1. Certified Value (mass fraction)

Sulfur: 8.41 mg/kg \pm 0.12 mg/kg

Information Values: Information values are provided in Table 2 for additional properties of SRM 1616b. The values are not certified and are given to provide additional information on the matrix. A NIST information value is a value that may be of interest to the SRM user, but insufficient information is available to assess adequately the uncertainties associated with the values [3]. Information values cannot be used to establish metrological traceability.

Expiration of Certification: The certification of **SRM 1616b** is valid, within the measurement uncertainty specified, until **31 December 2018**, provided the SRM is handled and stored in accordance with instructions given in this certificate (see "Instructions for Handling, Storage, and Use"). The certification is nullified if the SRM is damaged, contaminated, or otherwise modified.

Maintenance of SRM Certification: NIST will monitor this SRM over the period of its certification. If substantive technical changes occur that affect the certification before the expiration of this certificate, NIST will notify the purchaser. Registration (see attached sheet or register online) will facilitate notification.

Coordination of the technical measurements leading to the certification of this SRM was provided by W.R. Kelly and G.C. Turk of the NIST Chemical Sciences Division.

Analytical measurements by ID-TIMS for certification were performed by W.R. Kelly, J.L. Mann, and R.D. Vocke and homogeneity testing by X-ray fluorescence spectrometry was performed by A.F. Marlow and J.R. Sieber of the NIST Chemical Sciences Division.

Statistical analysis for this SRM was provided by W.F. Guthrie of the NIST Statistical Engineering Division.

Carlos A. Gonzalez, Chief
Chemical Sciences Division

Robert L. Watters, Jr., Director
Office of Reference Materials

Gaithersburg, MD 20899
Certificate Issue Date: 30 November 2015
Certificate Revision History on Last Page

Bottling was performed under the supervision of M.P. Cronise of the NIST Office of Reference Materials.

The kerosene for this SRM was donated by BP Products North America, Naperville, IL.⁽¹⁾

Support aspects involved in the issuance of this SRM were coordinated through the NIST Office of Reference Materials.

INSTRUCTIONS FOR HANDLING, STORAGE, AND USE

Each SRM bottle should only be opened for the minimum time required to dispense the material. To relate analytical determinations to the certified value in this Certificate of Analysis, a minimum sample mass of 150 mg should be used. The tightly capped bottle should be stored under normal laboratory conditions away from direct sunlight.

Table 2. Information Values for Selected Properties

Physical Property Test ^(a)	ASTM Standard Used	Result
Density @ 15 °C @ 60 °F	D 1250-80 (1990) ^{ε1} D 4052-96	827.1 kg/m ³ 39.5 API
Flash Point	D 93 (A)-94	58.9 °C
Kinematic Viscosity @ 40 °C	D 445-94 ^{ε1}	1.617 × 10 ⁻⁶ m ² /s (1.617 cSt)
Carbon	D 5291-92	84.0 %
Hydrogen	D 5291-92	13.1 %

^(a) These properties were determined by a commercial firm under contract to NIST using ASTM methods. The results are **NOT** certified and are provided as additional information on the matrix.

ASTM Standards

D 93-94	Standard Test Methods for Flash Point by Pensky-Martens Closed Cup Tester
D 4052-96	Standard Test Method for Density and Relative Density of Liquids by Digital Density Meter
D 445-94 ^{ε1}	Standard Test Method for Kinematic Viscosity of Transparent and Opaque Liquids (the Calculation of Dynamic Viscosity)
D 1250-80 (1990) ^{ε1}	Standard Guide for Petroleum Measurement Tables
D 2274-94	Standard Test Method for Oxidation Stability of Distillate Fuel Oil (Accelerated Method)
D 5291-92	Standard Test Methods for Instrumental Determination of Carbon, Hydrogen, and Nitrogen in Petroleum Products and Lubricants

^{ε1}: A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval. The superscript number refers to the specific change.

⁽¹⁾Certain commercial equipment, instruments or materials are identified in this certificate to adequately specify the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.

REFERENCES

- [1] ASTM D 3699-92; *Standard Specification for Diesel Fuel Oils*; Annual Book of ASTM Standards, Vol. 05.01 (1998).
- [2] Kelly, W.R.; Paulsen, P.J.; Murphy, K.E.; Vocke, R.D., Jr.; Chen, L.-T.; *Determination of Sulfur in Fossil Fuels by Isotope Dilution Thermal Ionization Mass Spectrometry*; Anal. Chem., Vol. 66, pp. 2505–2513 (1994).
- [3] May, W.; Parris, R.; Beck, C.; Fassett, J.; Greenberg, R.; Guenther, F.; Kramer, G.; Wise, S.; Gills, T.; Colbert, J.; Gettings, R.; MacDonald, B.; *Definitions of Terms and Modes Used at NIST for Value-Assignment of Reference Materials for Chemical Measurements*; NIST Special Publication 260-136; U.S. Government Printing Office: Washington, DC (2000); available at: <http://www.nist.gov/srm/publications.cfm> (accessed Nov 2015).
- [4] JCGM 100:2008; *Evaluation of Measurement Data - Guide to the Expression of Uncertainty in Measurement*; (GUM 1995 with Minor Corrections), Joint Committee for Guides in Metrology (JCGM) (2008); available at http://www.bipm.org/utis/common/documents/jcgm/JCGM_100_2008_E.pdf (accessed Nov 2015); see also Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297, U.S. Government Printing Office: Washington, DC (1994); available at <http://www.nist.gov/pml/pubs/index.cfm> (accessed Nov 2015).

Certificate Revision History: **30 November 2015** (Change of expiration date; editorial changes); **31 August 2010** (Extension of the certification period; editorial changes); **14 February 2008** (Update of expiration date); **12 September 2006** (Editorial changes); **10 March 2005** (Original certificate date).

Users of this SRM should ensure that the Certificate of Analysis in their possession is current. This can be accomplished by contacting the SRM Program: telephone (301) 975-2200; fax (301) 948-3730, email srminfo@nist.gov; or via the Internet at <http://www.nist.gov/srm>.