

National Bureau of Standards Certificate

Standard Reference Material 44e

Aluminum

Secondary Freezing Point Standard

(International Practical Temperature Scale – 1968)

660.3 °C

The freezing point given above is that of a representative sample from lot 44e. It is a secondary standard and is not necessarily the same as the freezing point of pure aluminum. It is estimated that the uncertainty in the NBS certified value of freezing point does not exceed ± 0.2 degree. (The estimated uncertainty includes the imprecision and possible systematic errors of the measurements.)

Full details of the precautions that should be observed in freezing point determinations are given in National Bureau of Standards Circular 590. [For procedures employing purer metal samples, see Metrologia 7, No. 3, 108-130 (July 1971).] The International Practical Temperature Scale of 1968 is described in Metrologia 5, No. 2, 35-44 (April 1969). This is a revision of the certificate dated October 30, 1957, and consists primarily of the conversion of IPTS (1948) to IPTS (1968) by G. T. Furukawa.

Washington, D. C. 20234
December 6, 1971

J. Paul Cali, Chief
Office of Standard Reference Materials