



Report of Investigation

Reference Material 8550

USGS25

(Nitrogen Isotopes in Ammonium Sulfate)

This Reference Material (RM) is intended for use in developing and validating methods for measuring relative differences in nitrogen (N) isotope-number ratios, $R(^{15}\text{N}/^{14}\text{N})$ [1]. Even though the value for this RM is a reference value and not certified [2], its use will improve the comparability of data from different laboratories. The equivalent name used by the International Atomic Energy Agency (IAEA) and the U.S. Geological Survey (USGS) for this RM is USGS25. RM 8550 consists of one bottle containing approximately 0.4 g of ammonium sulfate salt $((\text{NH}_4)_2\text{SO}_4)$.

Table 1. Reference Value^(a) and Expanded Uncertainty for the Relative N Isotope-Number Ratio Difference of RM 8550.

RM Number	Name	Reference Value $10^3 \delta^{15}\text{N}_{\text{AIR}}^{(b)}$	Expanded Uncertainty $10^3 \delta^{15}\text{N}_{\text{AIR}}^{(b)}$
8550	USGS25	-30.41	± 0.16

^(a) A reference value is a non-certified value that is the best estimate of the true value; however, the value may reflect only the measurement precision and may not include all sources of uncertainty [2].

^(b) The $\delta^{15}\text{N}_{\text{AIR}}$ value is expressed as a mean and an expanded uncertainty. The expanded uncertainty is equal to $U = ku_c$, where u_c is the combined standard uncertainty as defined by the ISO Guide [3] and k is the coverage factor. The combined standard uncertainty is intended to represent, at the level of one standard deviation, the effects of random errors on the reference value that were evaluated by statistical means (Type A). Any uncertainty due to biases in the methods is not included in the expanded uncertainty. The coverage factor, $k = 2.179$ ($n=13$), provides an expanded uncertainty interval that has about a 95 % probability of encompassing the mean. The values for RM 8550 are taken from reference 4 (AIR – N₂ of tropospheric air).

Reference Difference in Isotope-Number Ratio Values: The differences in measured isotope-number ratios of stable nitrogen isotopes in substance P, $R(^{15}\text{N}/^{14}\text{N})_P = [N(^{15}\text{N})_P / N(^{14}\text{N})_P]$, are reported as $\delta^{15}\text{N}$ values [5]. The relative differences in isotope-number ratios for nitrogen are referenced to AIR where:

$$\delta^{15}\text{N} = [R(^{15}\text{N}/^{14}\text{N})_{\text{sample}} / R(^{15}\text{N}/^{14}\text{N})_{\text{AIR}}] - 1$$

AIR refers to N₂ of tropospheric air [6], for which $R(^{15}\text{N}/^{14}\text{N})_{\text{AIR}} = 0.003677$ [5,7,8]. The reported $\delta^{15}\text{N}$ results have been normalized to yield a value of +180 ‰ for RM 8558 (USGS32) [4], where the symbol ‰ is part per thousand and is equal to 0.001.

Expiration of Value Assignment: RM 8550 is valid, within the measurement uncertainty specified, until **31 December 2020**, provided the RM is handled in accordance with instructions given in this Report of Investigation (see “Instructions for Handling, Storage, and Use”). This report is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of RM Certification: NIST will monitor this RM over the period of its validity. If substantive technical changes occur that affect the value assignment before the expiration of this report, NIST will notify the purchaser. Registration (see attached sheet) will facilitate notification.

Support and technical aspects of the preparation, analysis, and distribution of this RM was coordinated through the NIST Chemical Sciences Division by R.D. Vocke, Jr.

Support aspects involved in the issuance of this RM were coordinated through the NIST Office of Reference Materials.

Carlos A. Gonzalez, Chief
Chemical Sciences Division

Gaithersburg, MD 20899
Report Issue Date: 30 January 2013
Report Revision History on Last Page

Robert L. Watters, Jr., Director
Office of Reference Materials

INSTRUCTIONS FOR HANDLING, STORAGE, AND USE

Handling and Storage: This RM should be kept in a dry environment as it will attract water when exposed to air. To minimize the potential for contamination, it is recommended that this RM be stored in the container in which it is supplied.

Use: It is recommended that RM 8550 (USGS25) be used to calibrate or check working laboratory standards that are used routinely. Working laboratory standards with a range of N isotope-ratios can be produced by the methods described by Böhlke *et al.* [9,10].

Distribution: The distribution of RM 8550 (USGS25) is limited to one unit per customer per three-year period of time.

PREPARATION AND ANALYSIS

Preparation: RM 8550 (USGS25) was prepared as a dried salt by dissolving and recrystallizing a mixture of normal reagent salt and ¹⁵N-depleted salt by J.K. Böhlke (USGS, Reston, Virginia) [4].

Analytical Methods: The reference value for $\delta^{15}\text{N}$ in RM 8550 (USGS25) was derived from an inter-laboratory comparison test after elimination of outliers [4]. The $\delta^{15}\text{N}$ values were measured by mass spectrometry on N₂ gas that was quantitatively produced using variants of a buffered sample combustion method coupled with additional purification steps. The measured results were then normalized to yield a value of +180 ‰ for RM 8558 (USGS32).

The $\delta^{15}\text{N}$ value and expanded uncertainty reported in Table 1 are the values accepted by the Commission on Isotopic Abundances and Atomic Weights of the International Union of Pure and Applied Chemistry (IUPAC) (<http://ciaaw.org/Nitrogen.htm>) and the IAEA as of the date of this report.

Isotopic Homogeneity: There is no evidence of isotopic heterogeneity in this RM for sample sizes in the range of 10 μmol to 100 μmol of nitrogen [4].

Normalization of Data: The $\delta^{15}\text{N}$ values in samples should be normalized to the AIR-USGS32 scale by calibrating the measurement with respect to atmospheric N₂ [5] and the $\delta^{15}\text{N}$ value of the ¹⁵N-enriched anchor, RM 8558 (USGS32) [4]. A general formula for normalizing nitrogen isotope measurement results using two laboratory standards LS1 (AIR) and LS2 (USGS32) can be expressed as:

$$\delta^{15}\text{N}_{\text{sample,cal}} = \delta^{15}\text{N}_{\text{LS1,cal}} + (\delta^{15}\text{N}_{\text{sample,WS}} - \delta^{15}\text{N}_{\text{LS1,WS}}) \times f \quad (1)$$

where the normalization factor f is:

$$f = \frac{(\delta^{15}\text{N}_{\text{LS2,cal}} - \delta^{15}\text{N}_{\text{LS1,cal}})}{(\delta^{15}\text{N}_{\text{LS2,WS}} - \delta^{15}\text{N}_{\text{LS1,WS}})} \quad (2)$$

Note: In the above formulas, cal denotes calibrated measurements made versus the AIR-USGS32 scale, and $\delta^{15}\text{N}_{\text{LS1,cal}}$ and $\delta^{15}\text{N}_{\text{LS2,cal}}$ are the conventionally fixed $\delta^{15}\text{N}$ values for AIR and USGS32. WS denotes measurements made versus a transfer gas (working standard), and $\delta^{15}\text{N}_{\text{LS1,WS}}$ and $\delta^{15}\text{N}_{\text{LS2,WS}}$ are the $\delta^{15}\text{N}$ values for calibrated laboratory working standards.

Reporting of Nitrogen Stable Isotope δ -values: The following recommendations from IUPAC are provided for reporting $\delta^{15}\text{N}$ values [5]. It is recommended that:

- the value of 272 be employed for ¹⁴N/¹⁵N of N₂ in air for the calculation of atom percent ¹⁵N from measured $\delta^{15}\text{N}$ values;
- all $\delta^{15}\text{N}$ values should be reported with respect to air (atmospheric nitrogen gas) and normalized to RM 8558 (USGS32).

In addition, researchers are encouraged to report the isotopic composition of RM 8550 (USGS25) and other internationally distributed nitrogen isotopic reference materials [11] in their publications, as appropriate to the method, as though they have been interspersed among unknowns.

Current Reports of Investigation (ROI) for all light stable isotopic Reference Materials mentioned in this report are available on the SRM web site [12].

REFERENCES

- [1] Coplen, T.B.; *Guidelines and Recommended Terms for Expression of Stable-Isotope-Ratio and Gas-Ratio Measurement Results*; Rapid Communications in Mass Spectrometry, Vol. 025, pp. 2538–2560 (2011); available at <http://onlinelibrary.wiley.com/doi/10.1002/rcm.5129/pdf> (accessed Jan 2013).
- [2] May, W.E.; Parris, R.M.; Beck II, C.M.; Fassett, J.D.; Greenberg, R.R.; Guenther, F.R.; Kramer, G.W.; Wise, S.A.; Gills, T.E.; Colbert, J.C.; Gettings, R.J.; MacDonald, B.S.; *Definitions of Terms and Modes Used at NIST for Value-Assignment of Reference Materials for Chemical Measurements*; NIST Spec. Pub. 260-136; U.S. Government Printing Office: Washington, DC (2000); available at <http://www.nist.gov/srm/publications.cfm> (accessed Jan 2013).
- [3] JCGM 100:2008; *Evaluation of Measurement Data — Guide to the Expression of Uncertainty in Measurement* (ISO GUM 1995 with Minor Corrections); Joint Committee for Guides in Metrology (2008); available at http://www.bipm.org/utis/common/documents/jcgm/JCGM_100_2008_E.pdf (accessed Nov 2012); see also Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Tech. Note 1297; U.S. Government Printing Office: Washington, DC (1994); available at <http://www.nist.gov/physlab/pubs/index.cfm> (accessed Jan 2013).
- [4] Böhlke, J.K.; Coplen, T.B.; *Interlaboratory Comparison of Reference Materials for Nitrogen-Isotope-Ratio Measurements, in Reference and Intercomparison Materials for Stable Isotopes of Light Elements*; In Proceedings of a consultants meeting: Vienna, Dec. 1993; IAEA-TECDOC-825; pp. 51–66 (1995); available at http://www-pub.iaea.org/MTCD/publications/PDF/te_825_prn.pdf (accessed Jan 2013).
- [5] Coplen, T.B.; Krouse, H.R.; Böhlke, J.K.; *Reporting of Nitrogen-Isotope Abundances*; Pure and Applied Chemistry, Vol. 64. pp. 907–908 (1992); available at <http://www.iupac.org/publications/pac/1992/pdf/6406x0907.pdf> (accessed Jan 2013).
- [6] Mariotti, A.; *Atmospheric nitrogen is a reliable standard for natural ¹⁵N abundance measurements*; Nature, Vol. 303, pp. 685–687 (1983).
- [7] Junk, G.; Svec, H.J.; *The Absolute Abundance of the Nitrogen Isotopes in the Atmosphere and Compressed Gas from Various Sources*; Geochimica Cosmochimica Acta, Vol. 14, pp. 234–243 (1958).
- [8] DeBievre, P.; Valkiers, S.; Peiser, H.S.; Taylor, P.D.P.; Hansen, P.; *Mass Spectrometric Methods for Determining Isotopic Composition and Molar Mass Traceable to the SI, Exemplified by Improved Values for Nitrogen*; Metrologia, Vol. 33, pp. 447–455 (1996).
- [9] Böhlke, J.K.; Mroczkowski, S.J.; Coplen, T.B.; *Oxygen Isotopes in Nitrate: New Reference Materials for ¹⁸O:¹⁷O:¹⁶O Measurements and Observations on Nitrate-Water Equilibration*; Rapid Communications in Mass Spectrometry, Vol. 17, pp. 1835–1846 (2003).
- [10] Böhlke, J.K.; Gwinn, C.J.; Coplen, T.B.; *New Reference Materials for Nitrogen-Isotope-Ratio Measurements*; Geostands Newsletter, Vol. 17, pp. 159–164 (1993).
- [11] Coplen, T.B.; Hopple, J.A.; Böhlke, J.K.; Peiser, H.S.; Rieder, S.E.; Krouse, H.R.; Rosman, K.J.R.; Ding, T.; Vocke, Jr., R.D.; Révész, K.M.; Lambert, A.; Taylor, P.; De Bièvre, P.; *Compilation of Minimum and Maximum Isotope Ratios of Selected Elements in Naturally Occurring Terrestrial Materials and Reagents*; U.S. Geological Survey Water-Resources Investigations Report 01-4222, p. 98 (2001); available at <http://pubs.usgs.gov/wri/wri014222/pdf/wri01-4222.pdf> (accessed Jan 2013).
- [12] *Light Stable Isotopic Materials (gas, liquid and solid forms)*; NIST SRM Order Request System; National Institute of Standards and Technology; U.S. Department of Commerce: Gaithersburg, MD 20899; available at <https://www-s.nist.gov/srmors/viewTableV.cfm?tableid=42> (accessed Jan 2013).

Report Revision History: 30 January 2013 (Uncertainty updated to an expanded uncertainty for $\delta^{15}\text{N}_{\text{AIR}}$; expiration date assigned; editorial changes); 03 February 1993 (Updated value for RM 8549; added RM 8558 to report); 22 June 1992 (Original report issue date).
--

Users of this RM should ensure that the Report of Investigation in their possession is current. This can be accomplished by contacting the SRM Program: telephone (301) 975-2200; fax (301) 948-3730; e-mail srminfo@nist.gov; or via the Internet at <http://www.nist.gov/srm>.